

Information on DELOLUXcontrol calibration

In order to ensure the consistent quality of your production process, we recommend a yearly calibration of your DELOLUXcontrol UV intensity meter. Of course, we will be pleased to perform this service for you. You may choose between the following basic procedures:

- Factory calibration (standard procedure)
- DAkkS accredited examination
- DAkkS accredited calibration

In cooperation with our calibration laboratory, we can offer you the following calibration services:

	Article number	Designation	Scope of services
DAKKS ACCR. TESTING	ZUDP19	DAkkS accr. test measuring head LED 1 mm / 9 mm	DAkkS accr. testing – Measuring head LED 1 mm or 9 mm incl. test certificate
	ZUDPUVBL	DAkkS accr. test measuring head UVA / Blue	DAkkS accr. testing – UVA or Blue measuring head incl. test certificate
DAKKS ACCR. CALIBRATION	ZUDKANz*	DAkkS accr. calibration DELOLUXcontrol Display unit	DAkkS accr. calibration of the direct current conversion in the display unit incl. calibration certificate
	ZUDK19	DAkkS accr. calibration measuring head LED 1 mm / 9 mm	DAkkS accr. calibration of the LED measuring heads 1 mm or 9 mm incl. calibration certificate
FACTORY CALIBRATION	ZUWKAnz	Factory calibration DELOLUXcontrol Display unit	Factory calibration of the display unit according to DC source incl. factory calibration certificate
	ZUWK19	Factory calibration measuring head LED 1 mm / 9 mm	Factory calibration measuring head LED 1 mm or 9 mm incl. factory calibration certificate
	ZUWKUVBL	Factory calibration measuring head UVA / Blue	Factory calibration measuring head UVA or Blue incl. factory calibration certificate

*ZUDKANz only possible for article numbers 9520108 and 9520340

For further information on these calibration and testing procedures, please refer to the summary of our laboratory mentioned above: [Calibration information](#)

If no specific calibration order in the form of a written order is placed with your shipment, it cannot be sent to the calibration laboratory.

Please send us a written order with the type and scope of the desired calibration assignment at the latest when you dispatch the device. Please note that a late order can lead to a delay of up to four weeks in completion.

To avoid unnecessary waiting times, your device should be at our premises by one of the arrival dates listed below at the latest. Therefore, please send it to:

► **DELO Industrial Adhesives**
 Equipment Service Team
 DELO-Allee 1
 86949 Windach
 Germany

Arrival date at DELO 2025

JANUARY	January 7	January 20	
FEBRUARY	February 3	February 17	
MARCH	March 3	March 17	March 31
APRIL	April 14	April 28	
MAY	May 12	May 26	
JUNE	June 9	June 23	
JULY	July 7	July 21	
AUGUST	August 4	August 18	
SEPTEMBER	September 1	September 15	September 29
OCTOBER	October 13	October 27	
NOVEMBER	November 19	November 24	
DECEMBER	December 08	December 18	



Devices that arrive after the deadline can only be considered in the next calibration period. Please also note that, depending on the calibration volume, the run may take to 5 weeks (from the respective date)! Please take this into account in your planning. If you have any technical questions, please feel free to contact our Equipment Service Team by calling +49 8193 9900-9612.



Of course, you can send your DELOLUXcontrol incl. detector heads to DELO at any time if repair is necessary.